
BOWA ACADEMY

Overview of electrosurgery-related symbols

Symbol	Designation
	Foot switch connection
	Neutral electrode insulated against earth when using HF
	Defibrillation-protected applied part of type CF
	Alternating current
	“ON”/“OFF” (push-button operation)
	During activation (of the HF device), HF energy in the radio frequency range 9 kHz to 400 GHz is applied, which generates electromagnetic radiation.
	Identification of electrical and electronic equipment according to directive 2002/96/EC (WEEE), see Disposal
	Identification of the (active) HF output; Caution: Dangerous electrical voltage.
	Caution: Sale and prescription restricted to physicians. For the attending physician only.
	Manufacturer
	Date of manufacture
	Consult instructions for use
	Equipotential
	Signal input of the fibre optic cable connection
	Signal output of the fibre optic cable connection
	Ethernet connection
	USB connection
	Audio (IN)
	UART communication interface
REF	Product number
SN	Serial number

Symbol	Designation
CE 0123	CE mark with number of notified body This product complies with the essential requirements of the Medical Device Regulation (EU) 2017/745.
	Conforms to Russian standards
IP X	Protection category
LOT	Lot
QTY	Quantity
	Use-by date
STERILE EO	Sterilised using ethylene oxide
	Single use only
	Do not resterilise
	Do not use if packaging is damaged
	Caution: Fragile!
	Keep away from direct sunlight
	Keep dry
	Caution
MD	Medical device
	Latex free
	Consult instructions for use
	Temperature limit
	Humidity range to which the device may be exposed
	Air pressure range to which the device may be exposed

Caution: This checklist does not replace the instructions for use. No claims against BOWA-electronic GmbH can be derived from the information and details provided therein. Depending on the individual circumstances, it may be necessary to deviate from the information given in this document. Please contact your BOWA MEDICAL representative for instructions on using the device.

© BOWA-electronic GmbH 2026

BOWA-electronic GmbH | Heinrich-Hertz-Str. 4 –10 | 72810 Gomaringen | Germany | Phone +49 7072 6002-0 | info@bowa-medical.com | www.bowa-medical.com